



TURKISH ACCREDITATION AGENCY

## ACCREDITATION CERTIFICATE

As a Calibration Laboratory

**PAKKENS YEDEK PARÇA VE MAKİNA SANAYİ VE TİCARET ANONİM ŞİRKETİ - Pakkens Kalibrasyon Laboratuvarı**

Central Address: MİNARELİÇAVUŞ BURSA OSB MAH. SARI CAD. No:50/ NİLÜFER/BURSA Bursa / Türkiye

is accredited in accordance with TS EN ISO/IEC 17025:2017 standard within the scope given in Annex following the assessment conducted by TURKAK.

**Accreditation Number : AB-0009-K**

**Accreditation Date : 09.03.2004**

**Revision Date / Number : 11.02.2025 / 19**


This certificate shall remain in force until **20.04.2028**, subject to continuing compliance with the standard **TS EN ISO/IEC 17025:2017**, related regulations and requirements.

Gülden Banu Müderrisoğlu  
Secretary General



Turkish Accreditation Agency (TURKAK) is a signatory to the European co-operation for Accreditation (EA) Multilateral Agreement (MLA) and International Laboratory Accreditation Cooperation (ILAC) Mutual Recognition Agreement (MRA) in the scope of ISO/IEC 17025.

*This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.*

 <p>Calibration TS EN ISO/IEC 17025 AB-0009-K</p>	<p><b>PAKKENS YEDEK PARÇA VE MAKİNA SANAYİ VE TİCARET ANONİM ŞİRKETİ - Pakkens Kalibrasyon Laboratuvarı</b></p> <p>Accreditation Nr: AB-0009-K Revision Nr: 19 Date: 11.02.2025</p> <p><b>Calibration Laboratory</b></p> <p><b>Address :</b> MINARELİÇAVUŞ BURSA OSB MAH. SARI CAD. No:50/ NİLÜFER/BURSA Bursa / Türkiye</p> <p><b>Phone :</b> +90 224 600 0200 <b>Fax :</b> - <b>Email :</b> aersoz@pakkens.com.tr <b>Website :</b> www.pakkens.com</p>
--	--

## Calibration and Measurement Capability (CMC)

## Pressure

Measured Quantity / Calibrated Items	Range	Measurement Conditions	Expanded Measurement Uncertainty (k=2)	Remarks / Calibration Method
<b>Relative Pressure</b>  Analog Manometer Digital Manometer Pressure Calibrator Difference Pressure Meter	$-95 \text{ kPa} \leq p \leq -4 \text{ kPa}$	Pneumatic	$41 \text{ Pa} + 3,6 \cdot 10^{-4} \cdot p$	The calibration is performed in accordance with the DKD-R 6-1 "Calibration of Pressure Gauges" guideline document.  $p$ : Pressure, (Pa)  (The "p" value in the formulas included in the measurement uncertainty statement indicates the relative pressure in Pa.)
<b>Relative Pressure</b>  Analog Manometer Digital Manometer Pressure Calibrator Difference Pressure Meter	$1,5 \text{ kPa} \leq p \leq 100 \text{ kPa}$	Pneumatic	$33 \text{ Pa} + 3,7 \cdot 10^{-4} \cdot p$	The calibration is performed in accordance with the DKD-R 6-1 "Calibration of Pressure Gauges" guideline document.  $p$ : Pressure, (Pa)  (The "p" value in the formulas included in the measurement uncertainty statement indicates the relative pressure in Pa.)
<b>Relative Pressure</b>  Analog Manometer Digital Manometer Pressure Calibrator Difference Pressure Meter	$0,02 \text{ MPa} \leq p \leq 2,5 \text{ MPa}$	Pneumatic	$2,8 \cdot 10^2 \text{ Pa} + 3,9 \cdot 10^{-4} \cdot p$	The calibration is performed in accordance with the DKD-R 6-1 "Calibration of Pressure Gauges" guideline document.  $p$ : Pressure, (Pa)  (The "p" value in the formulas included in the measurement uncertainty statement indicates the relative pressure in Pa.)
<b>Relative Pressure</b>  Analog Manometer Digital Manometer Pressure Calibrator Difference Pressure Meter	$0,2 \text{ MPa} \leq p \leq 7,5 \text{ MPa}$	Pneumatic	$9,4 \cdot 10^2 \text{ Pa} + 3,9 \cdot 10^{-4} \cdot p$	The calibration is performed in accordance with the DKD-R 6-1 "Calibration of Pressure Gauges" guideline document.  $p$ : Pressure, (Pa)  (The "p" value in the formulas included in the measurement uncertainty statement indicates the relative pressure in Pa.)
<b>Relative Pressure</b>  Analog Manometer Digital Manometer Pressure Calibrator Difference Pressure Meter	$1 \text{ MPa} \leq p \leq 60 \text{ MPa}$	Pneumatic	$1,4 \cdot 10^4 \text{ Pa} + 4,1 \cdot 10^{-4} \cdot p$	The calibration is performed in accordance with the DKD-R 6-1 "Calibration of Pressure Gauges" guideline document.  $p$ : Pressure, (Pa)  (The "p" value in the formulas included in the measurement uncertainty statement indicates the relative pressure in Pa.)

This document has been signed by Gülden Banu Müderrisoğlu with a secure electronic signature in accordance with the electronic signature law numbered 5070. Use the QR code to verify the e-signed document.

